

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU062018\
 Data File : VU024775.D
 Acq On : 20 Jun 2018 22:47
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 21 08:33:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.626	0.545	12.9	82	0.00
3 P	Chloromethane	0.639	0.604	5.5	93	0.00
4 C	Vinyl Chloride	0.659	0.622	5.6#	89	0.00
5 T	Bromomethane	0.319	0.387	-21.3#	110	0.00
6 T	Chloroethane	0.431	0.405	6.0	97	0.00
7 T	Trichlorofluoromethane	1.006	0.975	3.1	93	0.00
8 T	Diethyl Ether	0.390	0.367	5.9	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.624	0.551	11.7	83	0.00
10 T	Methyl Iodide	0.451	0.501	-11.1	89	0.00
11 T	Tert butyl alcohol	0.199	0.168	15.6	83	0.00
12 CM	1,1-Dichloroethene	0.569	0.519	8.8#	84	0.00
13 T	Acrolein	0.104	0.100	3.8	91	0.00
14 T	Allyl chloride	1.057	0.889	15.9	80	0.00
15 T	Acrylonitrile	0.386	0.378	2.1	88	0.00
16 T	Acetone	0.439	0.341	22.3#	75	0.00
17 T	Carbon Disulfide	1.823	1.547	15.1	79	0.00
18 T	Methyl Acetate	0.940	0.992	-5.5	98	0.00
19 T	Methyl tert-butyl Ether	2.140	1.984	7.3	86	0.00
20 T	Methylene Chloride	0.688	0.632	8.1	89	0.00
21 T	trans-1,2-Dichloroethene	0.627	0.572	8.8	86	0.00
22 T	Diisopropyl ether	2.096	2.001	4.5	89	0.00
23 T	Vinyl Acetate	1.877	1.764	6.0	86	0.00
24 P	1,1-Dichloroethane	1.205	1.124	6.7	86	0.00
25 T	2-Butanone	0.589	0.556	5.6	85	0.00
26 T	2,2-Dichloropropane	1.145	0.882	23.0#	72	0.00
27 T	cis-1,2-Dichloroethene	0.721	0.675	6.4	88	0.00
28 T	Bromochloromethane	0.531	0.465	12.4	75	0.00
29 T	Tetrahydrofuran	0.365	0.353	3.3	90	0.00
30 C	Chloroform	1.220	1.164	4.6#	89	0.00
31 T	Cyclohexane	1.323	1.035	21.8#	85	0.00
32 T	1,1,1-Trichloroethane	1.128	1.042	7.6	87	0.00
33 S	1,2-Dichloroethane-d4	0.816	0.694	15.0	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.415	0.367	11.6	78	0.00
36 T	1,1-Dichloropropene	0.588	0.540	8.2	84	0.00
37 T	Ethyl Acetate	0.640	0.670	-4.7	88	0.00
38 T	Carbon Tetrachloride	0.646	0.617	4.5	86	0.00
39 T	Methylcyclohexane	0.733	0.677	7.6	85	0.00
40 TM	Benzene	1.674	1.640	2.0	88	0.00
41 T	Methacrylonitrile	0.345	0.362	-4.9	88	0.00
42 TM	1,2-Dichloroethane	0.638	0.626	1.9	89	0.00
43 T	Isopropyl Acetate	1.107	1.070	3.3	88	0.00
44 TM	Trichloroethene	0.476	0.451	5.3	88	0.00
45 C	1,2-Dichloropropane	0.447	0.450	-0.7#	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU062018\
 Data File : VU024775.D
 Acq On : 20 Jun 2018 22:47
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 21 08:33:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.314	0.311	1.0	87	0.00
47 T	Bromodichloromethane	0.625	0.607	2.9	88	0.00
48 T	Methyl methacrylate	0.539	0.539	0.0	89	0.00
49 T	1,4-Dioxane	0.011	0.012	-9.1	106	0.00
50 S	Toluene-d8	1.514	1.339	11.6	78	0.00
51 T	4-Methyl-2-Pentanone	0.695	0.711	-2.3	92	0.00
52 CM	Toluene	1.077	1.075	0.2#	90	0.00
53 T	t-1,3-Dichloropropene	0.702	0.677	3.6	85	0.00
54 T	cis-1,3-Dichloropropene	0.748	0.697	6.8	84	0.00
55 T	1,1,2-Trichloroethane	0.438	0.442	-0.9	94	0.00
56 T	Ethyl methacrylate	0.733	0.740	-1.0	92	0.00
57 T	1,3-Dichloropropane	0.739	0.758	-2.6	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.281	0.300	-6.8	93	0.00
59 T	2-Hexanone	0.568	0.567	0.2	91	0.00
60 T	Dibromochloromethane	0.546	0.521	4.6	88	0.00
61 T	1,2-Dibromoethane	0.487	0.493	-1.2	93	0.00
62 S	4-Bromofluorobenzene	0.601	0.544	9.5	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.475	0.435	8.4	84	0.00
65 PM	Chlorobenzene	1.277	1.256	1.6	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.485	0.472	2.7	92	0.00
67 C	Ethyl Benzene	2.239	2.157	3.7#	90	0.00
68 T	m/p-Xylenes	0.860	0.847	1.5	92	0.00
69 T	o-Xylene	0.851	0.843	0.9	94	0.00
70 T	Styrene	1.398	1.379	1.4	92	0.00
71 P	Bromoform	0.461	0.431	6.5	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.839	3.725	3.0	94	0.00
74 T	N-amyl acetate	1.826	1.757	3.8	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.286	1.284	0.2	100	0.00
76 T	1,2,3-Trichloropropane	1.101	1.099	0.2	96	0.00
77 T	Bromobenzene	0.972	0.964	0.8	96	0.00
78 T	n-propylbenzene	4.522	4.258	5.8	90	0.00
79 T	2-Chlorotoluene	2.667	2.550	4.4	93	0.00
80 T	1,3,5-Trimethylbenzene	3.327	3.245	2.5	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.476	0.397	16.6	81	0.00
82 T	4-Chlorotoluene	3.104	2.963	4.5	92	0.00
83 T	tert-Butylbenzene	3.201	3.244	-1.3	99	0.00
84 T	1,2,4-Trimethylbenzene	3.391	3.326	1.9	95	0.00
85 T	sec-Butylbenzene	4.006	3.931	1.9	95	0.00
86 T	p-Isopropyltoluene	3.588	3.468	3.3	94	0.00
87 T	1,3-Dichlorobenzene	1.795	1.765	1.7	95	0.00
88 T	1,4-Dichlorobenzene	1.807	1.787	1.1	96	0.00
89 T	n-Butylbenzene	3.186	2.944	7.6	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU062018\
Data File : VU024775.D
Acq On : 20 Jun 2018 22:47
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Jun 21 08:33:39 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.700	0.634	9.4	90	0.00
91 T	1,2-Dichlorobenzene	1.798	1.842	-2.4	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.355	0.345	2.8	91	0.00
93 T	1,2,4-Trichlorobenzene	1.111	1.215	-9.4	94	0.00
94 T	Hexachlorobutadiene	0.625	0.615	1.6	94	0.00
95 T	Naphthalene	3.293	4.094	-24.3#	103	0.00
96 T	1,2,3-Trichlorobenzene	1.107	1.276	-15.3	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6