

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062024\
 Data File : VU059376.D
 Acq On : 20 Jun 2024 13:39
 Operator : MD/SY
 Sample : P2960-11
 Misc : 5.05g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHB04

Quant Time: Jun 21 05:58:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 21 05:55:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	238136	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	241722	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	116197	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	38928	27.302	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	54.600%#
7) Chloroethane-d5	1.872	69	22233	18.203	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	36.400%#
11) 1,1-Dichloroethene-d2	2.538	63	55564	19.743	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	39.480%#
21) 2-Butanone-d5	4.624	46	81596	60.715	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	60.710%
24) Chloroform-d	5.052	84	89778	31.630	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	63.260%#
26) 1,2-Dichloroethane-d4	5.692	65	59308	34.091	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	68.180%#
32) Benzene-d6	5.714	84	184338	33.227	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	66.460%#
36) 1,2-Dichloropropane-d6	6.682	67	61448	32.123	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	64.240%#
41) Toluene-d8	7.891	98	155052	32.922	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	65.840%#
43) trans-1,3-Dichloroprop...	8.177	79	23734	30.695	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.380%
47) 2-Hexanone-d5	8.640	63	53819	60.841	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	60.840%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	102089	31.534	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	63.060%#
66) 1,2-Dichlorobenzene-d4	12.190	152	64063	36.944	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	73.880%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062024\
Data File : VU059376.D
Acq On : 20 Jun 2024 13:39
Operator : MD/SY
Sample : P2960-11
Misc : 5.05g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHB04

Quant Time: Jun 21 05:58:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 21 05:55:12 2024
Response via : Initial Calibration

