

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062024\
 Data File : VU059373.D
 Acq On : 20 Jun 2024 12:32
 Operator : MD/SY
 Sample : P2897-10ME
 Misc : 6.12g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COSL6ME

Quant Time: Jun 21 05:57:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 21 05:55:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	239196	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	247415	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	121255	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.593	65	38564	26.927	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	53.860%#
7) Chloroethane-d5	1.872	69	21793	17.764	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	35.520%#
11) 1,1-Dichloroethene-d2	2.538	63	54925	19.429	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	38.860%#
21) 2-Butanone-d5	4.621	46	76431	56.619	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	56.620%
24) Chloroform-d	5.049	84	89478	31.384	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	62.760%#
26) 1,2-Dichloroethane-d4	5.692	65	57365	32.828	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	65.660%#
32) Benzene-d6	5.715	84	182786	32.189	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	64.380%#
36) 1,2-Dichloropropane-d6	6.682	67	58272	29.762	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	59.520%#
41) Toluene-d8	7.891	98	153529	31.849	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	63.700%#
43) trans-1,3-Dichloroprop...	8.174	79	24521	30.983	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.960%
47) 2-Hexanone-d5	8.637	63	51827	57.241	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	57.240%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	101854	30.737	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	61.480%#
66) 1,2-Dichlorobenzene-d4	12.190	152	67925	37.537	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	75.080%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062024\
Data File : VU059373.D
Acq On : 20 Jun 2024 12:32
Operator : MD/SY
Sample : P2897-10ME
Misc : 6.12g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0SL6ME

Quant Time: Jun 21 05:57:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 21 05:55:12 2024
Response via : Initial Calibration

