

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062424\
 Data File : VU059435.D
 Acq On : 24 Jun 2024 09:33
 Operator : MD/SY
 Sample : VU0624MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK083

Quant Time: Jun 24 12:44:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 22 00:01:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	222802	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	218825	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	92669	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	69721	52.264	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.520%
7) Chloroethane-d5	1.907	69	57009	49.888	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.780%
11) 1,1-Dichloroethene-d2	2.560	63	81960	31.126	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.260%
21) 2-Butanone-d5	4.611	46	92565	73.617	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	73.620%
24) Chloroform-d	5.052	84	122501	46.128	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.260%
26) 1,2-Dichloroethane-d4	5.692	65	76179	46.802	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.600%
32) Benzene-d6	5.721	84	249610	49.700	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.400%
36) 1,2-Dichloropropane-d6	6.682	67	80104	46.257	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.520%
41) Toluene-d8	7.891	98	212596	49.864	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.720%
43) trans-1,3-Dichloroprop...	8.174	79	32161	45.945	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.900%
47) 2-Hexanone-d5	8.627	63	54040	67.483	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	67.480%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	118978	40.596	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	77682	56.172	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.340%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062424\
Data File : VU059435.D
Acq On : 24 Jun 2024 09:33
Operator : MD/SY
Sample : VU0624MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK083

Quant Time: Jun 24 12:44:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 22 00:01:31 2024
Response via : Initial Calibration

