

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062524\
 Data File : VU059461.D
 Acq On : 25 Jun 2024 11:12
 Operator : MD/SY
 Sample : P2936-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B34

Quant Time: Jun 26 04:27:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 26 04:26:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	145909	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	152881	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	63803	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	39613	3.904	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.000%	
7) Chloroethane-d5	1.907	69	43902	4.497	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.000%	
11) 1,1-Dichloroethene-d2	2.560	65	16400	3.896	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	78.000%	
20) 2-Butanone-d5	4.618	46	126036	60.307	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	120.620%	
24) Chloroform-d	5.055	84	99163	4.762	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.200%	
26) 1,2-Dichloroethane-d4	5.695	65	52516	5.235	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.800%	
32) Benzene-d6	5.721	84	190668	4.532	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.600%	
36) 1,2-Dichloropropane-d6	6.685	67	65550	5.084	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.600%	
41) Toluene-d8	7.894	98	149783	3.829	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	76.600%	
43) trans-1,3-Dichloroprop...	8.180	79	10479	2.596	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	52.000%#	
46) 2-Hexanone-d5	8.627	63	84362	50.291	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	100.580%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	57731	5.419	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	108.400%	
66) 1,2-Dichlorobenzene-d4	12.190	152	62720	5.523	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	110.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062524\
Data File : VU059461.D
Acq On : 25 Jun 2024 11:12
Operator : MD/SY
Sample : P2936-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B34

Quant Time: Jun 26 04:27:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 26 04:26:20 2024
Response via : Initial Calibration

