

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062624\
 Data File : VU059529.D
 Acq On : 26 Jun 2024 17:15
 Operator : MD/SY
 Sample : P2909-11ME
 Misc : 4.37g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4CP0ME

Quant Time: Jun 26 17:42:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 22 00:01:31 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	285090	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	289646	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	154917	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.593	65	94691	55.474	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	110.940%
7) Chloroethane-d5	1.869	69	45632	31.208	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	62.420%#
11) 1,1-Dichloroethene-d2	2.538	63	122884	36.472	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.940%
21) 2-Butanone-d5	4.618	46	166880	103.722	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.720%
24) Chloroform-d	5.052	84	182836	53.806	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.620%
26) 1,2-Dichloroethane-d4	5.692	65	117010	56.181	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.360%
32) Benzene-d6	5.715	84	378701	56.966	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	113.940%
36) 1,2-Dichloropropane-d6	6.682	67	120292	52.480	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.960%
41) Toluene-d8	7.891	98	332496	58.918	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	117.840%
43) trans-1,3-Dichloroprop...	8.174	79	52370	56.523	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	113.040%
47) 2-Hexanone-d5	8.634	63	124626	117.576	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	117.580%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	216678	55.855	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	111.700%
66) 1,2-Dichlorobenzene-d4	12.190	152	143976	62.276	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	124.560%#
Target Compounds						
					Qvalue	
15) Methyl Acetate	2.946	43	23467	7.834	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062624\
Data File : VU059529.D
Acq On : 26 Jun 2024 17:15
Operator : MD/SY
Sample : P2909-11ME
Misc : 4.37g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4CP0ME

Quant Time: Jun 26 17:42:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 22 00:01:31 2024
Response via : Initial Calibration

