

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062624\
 Data File : VU059523.D
 Acq On : 26 Jun 2024 15:00
 Operator : MD/SY
 Sample : VU0626MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK085

Quant Time: Jun 27 04:10:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 27 04:10:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	198530	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	204539	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	88738	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	60370	50.787	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.580%
7) Chloroethane-d5	1.908	69	51815	50.886	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.780%
11) 1,1-Dichloroethene-d2	2.560	63	76976	32.807	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.620%
21) 2-Butanone-d5	4.615	46	89300	79.703	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.700%
24) Chloroform-d	5.055	84	115143	48.659	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.320%
26) 1,2-Dichloroethane-d4	5.695	65	71536	49.323	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.640%
32) Benzene-d6	5.721	84	232900	49.611	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.220%
36) 1,2-Dichloropropane-d6	6.686	67	74227	45.857	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.720%
41) Toluene-d8	7.895	98	196429	49.290	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.580%
43) trans-1,3-Dichloroprop...	8.177	79	31199	47.684	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.360%
47) 2-Hexanone-d5	8.628	63	55151	73.681	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.680%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	117868	43.026	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.060%
66) 1,2-Dichlorobenzene-d4	12.190	152	74547	56.293	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062624\
Data File : VU059523.D
Acq On : 26 Jun 2024 15:00
Operator : MD/SY
Sample : VU0626MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK085

Quant Time: Jun 27 04:10:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 27 04:10:00 2024
Response via : Initial Calibration

