

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062624\
 Data File : VU059535.D
 Acq On : 26 Jun 2024 19:39
 Operator : MD/SY
 Sample : P2987-15ME
 Misc : 4.58g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0TA4ME

Quant Time: Jun 27 04:53:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 27 04:10:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	259954	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	265986	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	124896	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	80944	52.005	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.020%
7) Chloroethane-d5	1.872	69	36355	27.267	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	54.540%#
11) 1,1-Dichloroethene-d2	2.534	63	108558	35.335	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.680%
21) 2-Butanone-d5	4.624	46	142125	96.878	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.880%
24) Chloroform-d	5.052	84	161151	52.010	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.020%
26) 1,2-Dichloroethane-d4	5.692	65	101526	53.460	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.920%
32) Benzene-d6	5.714	84	333945	54.702	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.400%
36) 1,2-Dichloropropane-d6	6.682	67	108203	51.405	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.800%
41) Toluene-d8	7.891	98	294816	56.888	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.780%
43) trans-1,3-Dichloroprop...	8.174	79	44330	52.101	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.200%
47) 2-Hexanone-d5	8.637	63	105371	108.253	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.250%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	185333	52.025	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.040%
66) 1,2-Dichlorobenzene-d4	12.190	152	117400	62.987	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	125.980%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062624\
Data File : VU059535.D
Acq On : 26 Jun 2024 19:39
Operator : MD/SY
Sample : P2987-15ME
Misc : 4.58g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0TA4ME

Quant Time: Jun 27 04:53:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 27 04:10:00 2024
Response via : Initial Calibration

