

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062724\
 Data File : VU059552.D
 Acq On : 27 Jun 2024 10:56
 Operator : MD/SY
 Sample : P3062-01 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Jun 28 04:07:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 27 04:10:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	193644	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	197345	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	83832	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	56477	48.711	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	97.420%		
7) Chloroethane-d5	1.908	69	49800	50.142	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	100.280%		
11) 1,1-Dichloroethene-d2	2.560	63	70788	30.931	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	61.860%		
21) 2-Butanone-d5	4.612	46	90122	82.466	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	82.470%		
24) Chloroform-d	5.055	84	114160	49.460	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.920%		
26) 1,2-Dichloroethane-d4	5.695	65	72704	51.393	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	102.780%		
32) Benzene-d6	5.721	84	223708	49.390	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.780%		
36) 1,2-Dichloropropane-d6	6.682	67	72987	46.735	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	93.480%		
41) Toluene-d8	7.894	98	189319	49.237	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	98.480%		
43) trans-1,3-Dichloroprop...	8.174	79	30127	47.724	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	95.440%		
47) 2-Hexanone-d5	8.628	63	54720	75.770	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	75.770%		
56) 1,1,2,2-Tetrachloroeth...	10.750	84	116856	44.212	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	88.420%		
66) 1,2-Dichlorobenzene-d4	12.190	152	73467	58.724	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	117.440%		
Target Compounds						Qvalue
37) 1,2-Dichloropropane	6.682	63	7586	3.749	ug/L #	92
60) 1,2,3-Trichloropropane	11.807	75	9122	3.913	ug/L #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062724\
Data File : VU059552.D
Acq On : 27 Jun 2024 10:56
Operator : MD/SY
Sample : P3062-01 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Jun 28 04:07:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 27 04:10:00 2024
Response via : Initial Calibration

