

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070224\
 Data File : VU059731.D
 Acq On : 02 Jul 2024 12:05
 Operator : MD/SY
 Sample : VV0702WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Jul 03 01:46:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 03 01:39:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	297454	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	294111	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	132774	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	99195	51.938	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 103.880%		
7) Chloroethane-d5	1.908	69	82999	51.667	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 103.340%		
11) 1,1-Dichloroethene-d2	2.560	63	131662	38.804	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 77.600%		
21) 2-Butanone-d5	4.612	46	153812	115.886	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 115.890%		
24) Chloroform-d	5.052	84	185159	50.064	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.120%		
26) 1,2-Dichloroethane-d4	5.692	65	117814	53.574	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 107.140%		
32) Benzene-d6	5.718	84	393973	52.311	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.620%		
36) 1,2-Dichloropropane-d6	6.682	67	123754	53.092	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 106.180%		
41) Toluene-d8	7.891	98	352353	49.197	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 98.400%		
43) trans-1,3-Dichloroprop...	8.174	79	54274	49.759	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 99.520%		
47) 2-Hexanone-d5	8.624	63	110890	112.883	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 112.880%		
56) 1,1,2,2-Tetrachloroeth...	10.750	84	186390	52.098	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 104.200%		
66) 1,2-Dichlorobenzene-d4	12.187	152	131158	53.757	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 107.520%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070224\
Data File : VU059731.D
Acq On : 02 Jul 2024 12:05
Operator : MD/SY
Sample : VV0702WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Jul 03 01:46:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 03 01:39:57 2024
Response via : Initial Calibration

