

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070224\
 Data File : VU059730.D
 Acq On : 02 Jul 2024 11:42
 Operator : MD/SY
 Sample : VU0702MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK089

Quant Time: Jul 03 01:46:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 03 01:39:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	309810	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	298653	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	132538	50.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	102349	51.452	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.900%
7) Chloroethane-d5	1.908	69	85672	51.204	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.400%
11) 1,1-Dichloroethene-d2	2.560	63	134096	37.946	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.900%
21) 2-Butanone-d5	4.608	46	158094	114.361	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	114.360%
24) Chloroform-d	5.052	84	188710	48.989	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.980%
26) 1,2-Dichloroethane-d4	5.692	65	118276	51.639	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.280%
32) Benzene-d6	5.718	84	406001	53.088	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.180%
36) 1,2-Dichloropropane-d6	6.682	67	125451	53.002	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.000%
41) Toluene-d8	7.891	98	363195	49.939	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.880%
43) trans-1,3-Dichloroprop...	8.174	79	53868	48.636	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.280%
47) 2-Hexanone-d5	8.624	63	111735	112.013	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.010%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	188560	51.903	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	132470	54.391	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070224\
Data File : VU059730.D
Acq On : 02 Jul 2024 11:42
Operator : MD/SY
Sample : VU0702MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK089

Quant Time: Jul 03 01:46:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 03 01:39:57 2024
Response via : Initial Calibration

