

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070224\
 Data File : VU059739.D
 Acq On : 02 Jul 2024 15:06
 Operator : MD/SY
 Sample : P3108-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0ST8

Quant Time: Jul 03 01:55:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 03 01:39:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	266705	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	265951	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	119496	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	86941	50.771	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 101.540%		
7) Chloroethane-d5	1.907	69	73713	51.176	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 102.360%		
11) 1,1-Dichloroethene-d2	2.560	63	111863	36.770	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 73.540%		
21) 2-Butanone-d5	4.612	46	131336	110.360	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 110.360%		
24) Chloroform-d	5.052	84	162399	48.972	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 97.940%		
26) 1,2-Dichloroethane-d4	5.692	65	100884	51.165	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.320%		
32) Benzene-d6	5.718	84	338804	49.749	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 99.500%		
36) 1,2-Dichloropropane-d6	6.682	67	105617	50.109	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 100.220%		
41) Toluene-d8	7.891	98	299591	46.259	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 92.520%		
43) trans-1,3-Dichloroprop...	8.174	79	47128	47.783	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 95.560%		
47) 2-Hexanone-d5	8.624	63	91880	103.434	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 103.430%		
56) 1,1,2,2-Tetrachloroeth...	10.750	84	164414	50.821	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 101.640%		
66) 1,2-Dichlorobenzene-d4	12.187	152	113805	51.827	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 103.660%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070224\
Data File : VU059739.D
Acq On : 02 Jul 2024 15:06
Operator : MD/SY
Sample : P3108-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0ST8

Quant Time: Jul 03 01:55:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 03 01:39:57 2024
Response via : Initial Calibration

