

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070224\
 Data File : VU059744.D
 Acq On : 02 Jul 2024 17:06
 Operator : MD/SY
 Sample : P3108-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jul 03 01:56:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 03 01:39:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	254333	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	253967	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	112054	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	85543	52.384	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 104.760%		
7) Chloroethane-d5	1.907	69	72357	52.679	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 105.360%		
11) 1,1-Dichloroethene-d2	2.560	63	107684	37.118	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 74.240%		
21) 2-Butanone-d5	4.612	46	121610	107.158	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 107.160%		
24) Chloroform-d	5.055	84	157240	49.723	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 99.440%		
26) 1,2-Dichloroethane-d4	5.695	65	97782	52.004	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.000%		
32) Benzene-d6	5.718	84	327791	50.403	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.800%		
36) 1,2-Dichloropropane-d6	6.682	67	105308	52.320	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 104.640%		
41) Toluene-d8	7.891	98	287621	46.507	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 93.020%		
43) trans-1,3-Dichloroprop...	8.174	79	43642	46.336	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 92.680%		
47) 2-Hexanone-d5	8.624	63	84936	100.129	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 100.130%		
56) 1,1,2,2-Tetrachloroeth...	10.750	84	157534	50.993	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 101.980%		
66) 1,2-Dichlorobenzene-d4	12.187	152	110185	53.511	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 107.020%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070224\
Data File : VU059744.D
Acq On : 02 Jul 2024 17:06
Operator : MD/SY
Sample : P3108-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jul 03 01:56:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 03 01:39:57 2024
Response via : Initial Calibration

