

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070524\
 Data File : VU059847.D
 Acq On : 05 Jul 2024 09:25
 Operator : MD/SY
 Sample : VU0705WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK014

Quant Time: Jul 06 03:00:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jul 06 02:59:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	160283	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	168703	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	71454	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	54452	4.886	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.800%
7) Chloroethane-d5	1.907	69	46936	4.377	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.600%
11) 1,1-Dichloroethene-d2	2.560	65	21196	4.584	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.600%
20) 2-Butanone-d5	4.615	46	112011	48.790	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.580%
24) Chloroform-d	5.055	84	106437	4.653	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.695	65	52537	4.768	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.400%
32) Benzene-d6	5.721	84	213237	4.593	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.800%
36) 1,2-Dichloropropane-d6	6.682	67	68401	4.808	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.200%
41) Toluene-d8	7.891	98	183475	4.251	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.000%
43) trans-1,3-Dichloroprop...	8.174	79	23784	5.339	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.800%
46) 2-Hexanone-d5	8.627	63	91449	49.403	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.800%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	56658	4.820	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	69944	5.500	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070524\
Data File : VU059847.D
Acq On : 05 Jul 2024 09:25
Operator : MD/SY
Sample : VU0705WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK014

Quant Time: Jul 06 03:00:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jul 06 02:59:52 2024
Response via : Initial Calibration

