

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070720\
 Data File : VU039319.D
 Acq On : 07 Jul 2020 17:15
 Operator : SY/MD
 Sample : L3193-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4276

Quant Time: Jul 08 01:54:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 08 01:44:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	110077	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	105849	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	50928	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	37628	5.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.00%
7) Chloroethane-d5	1.92	69	34798	5.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.60%
11) 1,1-Dichloroethene-d2	2.58	63	56465	4.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.00%
20) 2-Butanone-d5	4.66	46	127674	43.74	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.48%
24) Chloroform-d	5.08	84	77490	5.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.00%
26) 1,2-Dichloroethane-d4	5.72	65	47833	5.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.20%
32) Benzene-d6	5.74	84	154078	6.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	121.00%
36) 1,2-Dichloropropane-d6	6.71	67	52415	5.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.00%
41) Toluene-d8	7.91	98	134506	6.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	121.80%
43) trans-1,3-Dichloropropene-	8.19	79	21678	5.48	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	109.60%
46) 2-Hexanone-d5	8.64	63	132846	56.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.30%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	49088	5.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	53867	5.93	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.60%

Target Compounds

					Ovalue
5) Vinyl chloride	1.61	62	34448	3.838 ug/L	94
13) Acetone	2.66	43	3085	1.958 ug/L	71
14) Carbon disulfide	2.81	76	3002	0.130 ug/L #	89
18) trans-1,2-Dichloroethene	3.37	96	4773	0.659 ug/L	87
22) cis-1,2-Dichloroethene	4.69	96	41671	5.128 ug/L	98
33) Benzene	5.78	78	3312	0.105 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070720\
Data File : VU039319.D
Acq On : 07 Jul 2020 17:15
Operator : SY/MD
Sample : L3193-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4276

Quant Time: Jul 08 01:54:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 08 01:44:57 2020
Response via : Initial Calibration

