

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071222\
 Data File : VU049687.D
 Acq On : 12 Jul 2022 10:49
 Operator : SY/MD
 Sample : VU0712MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK150

Quant Time: Jul 13 06:52:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 13 06:51:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	212561	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	196451	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	81673	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	89894	50.123	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 100.240%		
7) Chloroethane-d5	1.906	69	65855	50.799	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 101.600%		
11) 1,1-Dichloroethene-d2	2.565	63	117535	37.101	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 74.200%		
21) 2-Butanone-d5	4.623	46	170647	111.459	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 111.460%		
24) Chloroform-d	5.063	84	157566	49.603	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 99.200%		
26) 1,2-Dichloroethane-d4	5.700	65	112384	52.275	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.540%		
32) Benzene-d6	5.726	84	309123	51.796	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 103.600%		
36) 1,2-Dichloropropane-d6	6.690	67	107531	53.567	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 107.140%		
41) Toluene-d8	7.896	98	258507	49.738	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 99.480%		
43) trans-1,3-Dichloroprop...	8.179	79	42951	50.783	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 101.560%		
47) 2-Hexanone-d5	8.632	63	95791	106.082	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 106.080%		
56) 1,1,2,2-Tetrachloroeth...	10.754	84	145458	51.786	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 103.580%		
66) 1,2-Dichlorobenzene-d4	12.192	152	84675	54.987	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 109.980%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071222\
Data File : VU049687.D
Acq On : 12 Jul 2022 10:49
Operator : SY/MD
Sample : VU0712MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK150

Quant Time: Jul 13 06:52:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 13 06:51:38 2022
Response via : Initial Calibration

