

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU072020\
 Data File : VU039572.D
 Acq On : 20 Jul 2020 16:01
 Operator : SY/MD
 Sample : L3346-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4C32

Quant Time: Jul 21 07:15:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 21 06:02:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	100468	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	98700	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	45056	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	21110	3.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.60%
7) Chloroethane-d5	1.92	69	27014	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.80%
11) 1,1-Dichloroethene-d2	2.58	63	34412	3.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.00%
20) 2-Butanone-d5	4.65	46	120249	45.13	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.26%
24) Chloroform-d	5.09	84	58764	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
26) 1,2-Dichloroethane-d4	5.72	65	39283	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
32) Benzene-d6	5.75	84	111818	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.71	67	38718	4.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.80%
41) Toluene-d8	7.91	98	96204	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
43) trans-1,3-Dichloropropene-	8.19	79	14674	3.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.60%
46) 2-Hexanone-d5	8.64	63	87604	40.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.12%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	34460	4.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	39239	4.88	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU072020\
Data File : VU039572.D
Acq On : 20 Jul 2020 16:01
Operator : SY/MD
Sample : L3346-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4C32

Quant Time: Jul 21 07:15:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 21 06:02:44 2020
Response via : Initial Calibration

