

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072424\
 Data File : VU060154.D
 Acq On : 24 Jul 2024 15:27
 Operator : MD/SY
 Sample : P3244-07DL 20X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YDZF2DL

Quant Time: Jul 25 05:14:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jul 25 05:11:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	159178	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	162850	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	71394	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	53997	4.876	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.600%
7) Chloroethane-d5	1.908	69	49928	4.853	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.000%
11) 1,1-Dichloroethene-d2	2.551	65	21774	5.036	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.800%
20) 2-Butanone-d5	4.631	46	143004	55.229	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.460%
24) Chloroform-d	5.052	84	101127	4.345	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
26) 1,2-Dichloroethane-d4	5.692	65	51949	4.793	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.800%
32) Benzene-d6	5.718	84	209267	4.796	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
36) 1,2-Dichloropropane-d6	6.682	67	67640	4.865	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.400%
41) Toluene-d8	7.891	98	172424	4.374	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	8.178	79	21543	4.552	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.000%
46) 2-Hexanone-d5	8.634	63	107694	51.279	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.560%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	58369	4.717	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	64740	5.024	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.400%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.599	62	43611	2.676	ug/L	93
16) Methylene chloride	3.036	84	14391	0.966	ug/L	90
22) cis-1,2-Dichloroethene	4.654	96	55026	3.887	ug/L	95
51) Chlorobenzene	9.444	112	16056	0.429	ug/L	98
65) 1,4-Dichlorobenzene	11.833	146	12818	0.501	ug/L	98
67) 1,2-Dichlorobenzene	12.210	146	37708	1.568	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072424\
Data File : VU060154.D
Acq On : 24 Jul 2024 15:27
Operator : MD/SY
Sample : P3244-07DL 20X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YDZF2DL

Quant Time: Jul 25 05:14:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jul 25 05:11:47 2024
Response via : Initial Calibration

