

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080624\
 Data File : VU060320.D
 Acq On : 06 Aug 2024 14:07
 Operator : MD/SY
 Sample : P3204-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-EDBCP-WS

Quant Time: Aug 07 07:01:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Aug 06 11:07:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	30707	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.634	95	8760	0.837	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.000%	
68) 1,2-Dichlorobenzene-d4	12.193	152	10442	0.919	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.000%	
Target Compounds						
						Qvalue
56) 1,2-Dibromoethane	8.923	107	12483	1.826	ug/l	91
71) 1,2,3-Trichloropropane	10.817	75	15227	2.026	ug/l	87
86) 1,2-Dibromo-3-Chloropr...	13.000	75	647	0.440	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080624\
Data File : VU060320.D
Acq On : 06 Aug 2024 14:07
Operator : MD/SY
Sample : P3204-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-EDBCP-WS

Quant Time: Aug 07 07:01:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Aug 06 11:07:28 2024
Response via : Initial Calibration

