

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080724\
 Data File : VU060335.D
 Acq On : 07 Aug 2024 14:26
 Operator : MD/SY
 Sample : P3204-07DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-RVOA-WSDL

Quant Time: Aug 08 05:57:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Aug 06 11:07:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	29887	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.631	95	10510	1.032	ug/l	0.00
Spiked Amount 1.000			Recovery	=	103.000%	
68) 1,2-Dichlorobenzene-d4	12.190	152	11002	0.995	ug/l	0.00
Spiked Amount 1.000			Recovery	=	99.000%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.602	62	82818	6.303	ug/l	98
9) 1,1-Dichloroethene	2.576	96	26966	3.018	ug/l	90
15) Methylene Chloride	3.039	84	48887	4.371	ug/l	92
16) trans-1,2-Dichloroethene	3.351	96	11079	1.157	ug/l	91
22) cis-1,2-Dichloroethene	4.663	96	16641	1.623	ug/l	86
28) 1,1,1-Trichloroethane	5.309	97	16790	1.122	ug/l	84
30) Carbon Tetrachloride	5.518	117	14030	1.151	ug/l	94
35) Benzene	5.769	78	39227	1.062	ug/l	100
36) 1,2-Dichloroethane	5.792	62	15033	1.464	ug/l	98
37) Trichloroethene	6.537	130	4885	0.534	ug/l	95
38) 1,2-Dichloropropane	6.785	63	14748	1.476	ug/l	99
49) Toluene	7.965	92	49341	2.409	ug/l	97
52) 1,1,2-Trichloroethane	8.396	97	18627	2.510	ug/l	99
58) Tetrachloroethene	8.547	164	28603	3.500	ug/l	98
59) Chlorobenzene	9.444	112	23520	1.061	ug/l	100
63) Ethyl Benzene	9.566	91	131355	3.417	ug/l	99
64) m/p-Xylenes	9.692	106	9847	1.446	ug/l	90
65) o-Xylene	10.097	106	8773	0.961	ug/l	96
66) Styrene	10.110	104	74280	3.016	ug/l	93
83) 1,4-Dichlorobenzene	11.833	146	32673	1.732	ug/l	99
85) 1,2-Dichlorobenzene	12.209	146	28512	1.551	ug/l	98
87) 1,2,4-Trichlorobenzene	13.836	180	22724	2.162	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080724\
Data File : VU060335.D
Acq On : 07 Aug 2024 14:26
Operator : MD/SY
Sample : P3204-07DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-RVOA-WSDL

Quant Time: Aug 08 05:57:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Aug 06 11:07:28 2024
Response via : Initial Calibration

