

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081522\
 Data File : VU050342.D
 Acq On : 15 Aug 2022 15:34
 Operator : SY/MD
 Sample : N4160-21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Aug 16 00:32:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 16 00:29:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	98362	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	97791	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	36862	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	32805	4.008	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.200%
7) Chloroethane-d5	1.909	69	29470	4.462	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.200%
11) 1,1-Dichloroethene-d2	2.565	65	11959	3.738	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.800%
20) 2-Butanone-d5	4.629	46	110622	55.417	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.840%
24) Chloroform-d	5.063	84	66884	4.393	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
26) 1,2-Dichloroethane-d4	5.703	65	36295	4.643	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.800%
32) Benzene-d6	5.729	84	119048	3.901	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.000%
36) 1,2-Dichloropropane-d6	6.690	67	44133	4.392	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.800%
41) Toluene-d8	7.899	98	85547	3.174	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.400%#
43) trans-1,3-Dichloroprop...	8.182	79	13347	3.706	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.200%
46) 2-Hexanone-d5	8.635	63	56151	43.970	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.940%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	41159	4.664	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.200%
66) 1,2-Dichlorobenzene-d4	12.195	152	31332	4.338	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.800%
Target Compounds						Qvalue
16) Methylene chloride	3.044	84	3043	0.293	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081522\
Data File : VU050342.D
Acq On : 15 Aug 2022 15:34
Operator : SY/MD
Sample : N4160-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Aug 16 00:32:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 16 00:29:27 2022
Response via : Initial Calibration

