

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081522\
 Data File : VU050346.D
 Acq On : 15 Aug 2022 17:11
 Operator : SY/MD
 Sample : N4072-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Aug 16 00:33:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 16 00:29:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	92791	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	94596	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	35688	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	30748	3.982	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.600%
7) Chloroethane-d5	1.913	69	27305	4.383	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.600%
11) 1,1-Dichloroethene-d2	2.562	65	11774	3.901	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.000%
20) 2-Butanone-d5	4.636	46	114808	60.967	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.940%
24) Chloroform-d	5.064	84	67437	4.695	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
26) 1,2-Dichloroethane-d4	5.703	65	37896	5.139	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.800%
32) Benzene-d6	5.726	84	124098	4.204	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
36) 1,2-Dichloropropane-d6	6.690	67	46026	4.735	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.600%
41) Toluene-d8	7.899	98	95608	3.667	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.400%
43) trans-1,3-Dichloroprop...	8.182	79	15028	4.314	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.200%
46) 2-Hexanone-d5	8.636	63	58204	47.117	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.240%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	43104	5.049	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.000%
66) 1,2-Dichlorobenzene-d4	12.192	152	35549	5.084	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.600%
Target Compounds						Qvalue
16) Methylene chloride	3.044	84	2120	0.216	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081522\
Data File : VU050346.D
Acq On : 15 Aug 2022 17:11
Operator : SY/MD
Sample : N4072-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Aug 16 00:33:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 16 00:29:27 2022
Response via : Initial Calibration

