

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081922\
 Data File : VU050383.D
 Acq On : 19 Aug 2022 22:19
 Operator : SY/MD
 Sample : N3734-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WS-NAPHTHALENE

Quant Time: Aug 20 06:26:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Aug 19 15:06:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.121	96	30325	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.639	95	8214	0.796	ug/l	0.00
Spiked Amount 1.000			Recovery	=	80.000%	
68) 1,2-Dichlorobenzene-d4	12.198	152	9678	0.793	ug/l	0.00
Spiked Amount 1.000			Recovery	=	79.000%	
Target Compounds						
14) Carbon Disulfide	2.790	76	362218	13.093	ug/l	99
15) Methylene Chloride	3.047	84	9395	0.777	ug/l	93
45) 4-Methyl-2-Pentanone	7.816	43	88811	16.335	ug/l	98
89) Naphthalene	14.092	128	753909m	24.184	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081922\
Data File : VU050383.D
Acq On : 19 Aug 2022 22:19
Operator : SY/MD
Sample : N3734-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WS-NAPHTHALENE

Quant Time: Aug 20 06:26:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Aug 19 15:06:59 2022
Response via : Initial Calibration

