

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082324\
 Data File : VU060531.D
 Acq On : 23 Aug 2024 23:59
 Operator : MD/SY
 Sample : P3711-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2499

Quant Time: Aug 24 01:51:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 24 01:46:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	82254	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	90195	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	36064	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	16646	4.582	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.600%
7) Chloroethane-d5	1.911	69	29950	5.441	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.800%
11) 1,1-Dichloroethene-d2	2.560	65	8978	4.193	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.800%
20) 2-Butanone-d5	4.634	46	60706	48.592	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.180%
24) Chloroform-d	5.055	84	57643	5.565	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.400%
26) 1,2-Dichloroethane-d4	5.695	65	31158	5.765	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.200%
32) Benzene-d6	5.721	84	96434	4.889	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.800%
36) 1,2-Dichloropropane-d6	6.682	67	33919	5.330	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.600%
41) Toluene-d8	7.891	98	82811	4.475	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
43) trans-1,3-Dichloroprop...	8.177	79	10350	4.545	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.000%
46) 2-Hexanone-d5	8.627	63	50916	46.701	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.400%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	26821	4.737	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	37579	7.067	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	141.400%#
Target Compounds						Qvalue
47) Tetrachloroethene	8.547	164	179152	28.187	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082324\
Data File : VU060531.D
Acq On : 23 Aug 2024 23:59
Operator : MD/SY
Sample : P3711-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2499

Quant Time: Aug 24 01:51:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 24 01:46:38 2024
Response via : Initial Calibration

