




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082324\
Data File : VU060521.D
Acq On : 23 Aug 2024 19:56
Operator : MD/SY
Sample : P3711-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 3

Instrument :
MSVOA_U
ClientSampleId :
E24B0

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/26/2024
Supervised By :Mahesh Dadoda 08/27/2024

Quant Time: Aug 24 01:49:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 24 01:46:38 2024
Response via : Initial Calibration



TIC: VU060521.D\data.ms

(13) Acetone (T)

2.447min (-0.203) 564.75 ug/L

response 661673

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	33.40	0.04
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082324\
Data File : VU060521.D
Acq On : 23 Aug 2024 19:56
Operator : MD/SY
Sample : P3711-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E24B0

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/26/2024
Supervised By :Mahesh Dadoda 08/27/2024

Quant Time: Aug 24 01:49:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 24 01:46:38 2024
Response via : Initial Calibration



TIC: VU060521.D\data.ms

(13) Acetone (T)

2.644min (-0.007) 20.55 ug/L m

response 24074

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	33.40	1.12
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082324\
 Data File : VU060521.D
 Acq On : 23 Aug 2024 19:56
 Operator : MD/SY
 Sample : P3711-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E24B0

Manual IntegrationsAPPROVED

Quant Time: Aug 24 01:49:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 24 01:46:38 2024
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 08/26/2024
 Supervised By :Mahesh Dadoda 08/27/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	6.242	114		92186	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117		115674	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152		39717	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.595	65		17067	4.192	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.800%	
7) Chloroethane-d5	1.911	69		30712	4.978	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.600%	
11) 1,1-Dichloroethene-d2	2.560	65		15577	6.491	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	129.800%#	
20) 2-Butanone-d5	4.631	46		72877	52.049	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.100%	
24) Chloroform-d	5.055	84		55881	4.814	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.200%	
26) 1,2-Dichloroethane-d4	5.695	65		31307	5.168	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.400%	
32) Benzene-d6	5.721	84		102741	4.062	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.200%	
36) 1,2-Dichloropropane-d6	6.682	67		38444	4.711	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.200%	
41) Toluene-d8	7.891	98		85338	3.596	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.000%	
43) trans-1,3-Dichloroprop...	8.177	79		10972	3.757	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.200%	
46) 2-Hexanone-d5	8.627	63		46094	32.966	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	65.940%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84		26873	3.700	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.000%	
66) 1,2-Dichlorobenzene-d4	12.187	152		33972	5.801	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.000%	
Target Compounds							
13) Acetone	2.644	43		24074m	20.548	ug/L	
42) Toluene	7.965	91		5794	0.149	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082324\
Data File : VU060521.D
Acq On : 23 Aug 2024 19:56
Operator : MD/SY
Sample : P3711-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E24B0

Manual IntegrationsAPPROVED

Quant Time: Aug 24 01:49:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 24 01:46:38 2024
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 08/26/2024
Supervised By :Mahesh Dadoda 08/27/2024

