

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083119\
 Data File : VU034092.D
 Acq On : 30 Aug 2019 12:48
 Operator : JC/SP
 Sample : K4555-11
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleID :
 TW-11

Quant Time: Aug 31 03:54:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	474129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	755206	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	734314	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	378928	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	293838	48.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.74%	
35) Dibromofluoromethane	4.86	113	241780	46.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.32%	
50) Toluene-d8	7.55	98	963352	49.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.24%	
62) 4-Bromofluorobenzene	10.29	95	349321	45.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.58%	
Target Compounds						
16) Acetone	2.31	43	22687	7.319	ug/l	96
17) Carbon Disulfide	2.47	76	6484	0.432	ug/l #	95
95) Naphthalene	13.74	128	9061	2.159	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083119\
Data File : VU034092.D
Acq On : 30 Aug 2019 12:48
Operator : JC/SP
Sample : K4555-11
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TW-11

Quant Time: Aug 31 03:54:58 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

