

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083119\
 Data File : VU034105.D
 Acq On : 30 Aug 2019 17:50
 Operator : JC/SP
 Sample : K4620-16MEDL 10X
 Misc : 6.28g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S504-K2-9.5-10.0-082919MEDL

Quant Time: Aug 31 04:19:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	527140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	823789	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	786661	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	393787	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	305214	45.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.38%	
35) Dibromofluoromethane	4.86	113	253441	44.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.72%	
50) Toluene-d8	7.55	98	1030245	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
62) 4-Bromofluorobenzene	10.29	95	377348	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.20	96	105502	16.412	ug/l	89
44) Trichloroethene	6.17	130	22154	3.579	ug/l	98
95) Naphthalene	13.74	128	15441	2.396	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU083119\
Data File : VU034105.D
Acq On : 30 Aug 2019 17:50
Operator : JC/SP
Sample : K4620-16MEDL 10X
Misc : 6.28g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S504-K2-9.5-10.0-082919MEDL

Quant Time: Aug 31 04:19:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

