

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083119\
 Data File : VU034106.D
 Acq On : 30 Aug 2019 18:13
 Operator : JC/SP
 Sample : K4591-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SB-3-30-32

Quant Time: Aug 31 04:21:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	534728	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	847956	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	803766	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	407356	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	315432	46.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.08%	
35) Dibromofluoromethane	4.86	113	264439	44.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.92%	
50) Toluene-d8	7.55	98	1065755	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
62) 4-Bromofluorobenzene	10.29	95	385123	44.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.92%	
Target Compounds						
16) Acetone	2.31	43	36346	10.396	ug/l	97
20) Methylene Chloride	2.69	84	9410	1.422	ug/l	96
43) Isopropyl Acetate	5.53	43	122856	8.991	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU083119\
Data File : VU034106.D
Acq On : 30 Aug 2019 18:13
Operator : JC/SP
Sample : K4591-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SB-3-30-32

Quant Time: Aug 31 04:21:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

