

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090922\
 Data File : VU050703.D
 Acq On : 09 Sep 2022 16:41
 Operator : SY/MD
 Sample : N4482-13
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 10 01:21:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 10 01:20:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	324511	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	315926	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	122416	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	128854	46.424	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	92.840%		
7) Chloroethane-d5	1.906	69	90707	48.327	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	96.660%		
11) 1,1-Dichloroethene-d2	2.565	63	184919	37.701	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	75.400%		
21) 2-Butanone-d5	4.620	46	170570	83.996	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	84.000%		
24) Chloroform-d	5.063	84	257380	48.927	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.860%		
26) 1,2-Dichloroethane-d4	5.703	65	163214	50.773	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	101.540%		
32) Benzene-d6	5.729	84	517015	49.521	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.040%		
36) 1,2-Dichloropropane-d6	6.690	67	165167	48.185	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	96.360%		
41) Toluene-d8	7.899	98	440559	49.345	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	98.700%		
43) trans-1,3-Dichloroprop...	8.179	79	67452	47.559	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	95.120%		
47) 2-Hexanone-d5	8.632	63	68244	79.150	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	79.150%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	240556	46.375	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	92.760%		
66) 1,2-Dichlorobenzene-d4	12.195	152	141488	49.373	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	98.740%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090922\
Data File : VU050703.D
Acq On : 09 Sep 2022 16:41
Operator : SY/MD
Sample : N4482-13
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 10 01:21:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 10 01:20:12 2022
Response via : Initial Calibration

