

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091624\
 Data File : VU060787.D
 Acq On : 16 Sep 2024 18:13
 Operator : MD/SY
 Sample : P3918-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 17 01:56:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Sep 17 01:47:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	149858	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	161946	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	63672	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	63739	4.729	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.600%
7) Chloroethane-d5	1.911	69	47721	4.973	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.400%
11) 1,1-Dichloroethene-d2	2.563	65	22943	4.802	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.000%
20) 2-Butanone-d5	4.637	46	100288	46.347	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.700%
24) Chloroform-d	5.055	84	101154	4.909	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	5.695	65	50327	5.110	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%
32) Benzene-d6	5.721	84	204148	4.703	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
36) 1,2-Dichloropropane-d6	6.682	67	65995	4.914	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.200%
41) Toluene-d8	7.891	98	167328	4.228	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	8.174	79	22163	4.585	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.600%
46) 2-Hexanone-d5	8.631	63	85355	44.760	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.520%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	50630	4.812	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	56326	5.265	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091624\
Data File : VU060787.D
Acq On : 16 Sep 2024 18:13
Operator : MD/SY
Sample : P3918-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 17 01:56:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Sep 17 01:47:23 2024
Response via : Initial Calibration

