

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091624\
 Data File : VU060797.D
 Acq On : 16 Sep 2024 22:15
 Operator : MD/SY
 Sample : P3967-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COAN4

Quant Time: Sep 17 01:58:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Sep 17 01:47:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	137887	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	145990	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	59969	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	58487	4.716	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	94.400%	
7) Chloroethane-d5	1.911	69	45982	5.207	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.200%	
11) 1,1-Dichloroethene-d2	2.563	65	22094	5.026	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	100.600%	
20) 2-Butanone-d5	4.634	46	98739	49.593	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	99.180%	
24) Chloroform-d	5.055	84	95656	5.045	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.000%	
26) 1,2-Dichloroethane-d4	5.695	65	48164	5.315	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.200%	
32) Benzene-d6	5.721	84	193162	4.937	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.800%	
36) 1,2-Dichloropropane-d6	6.682	67	61941	5.116	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.400%	
41) Toluene-d8	7.891	98	161136	4.517	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.400%	
43) trans-1,3-Dichloroprop...	8.174	79	21685	4.976	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	99.600%	
46) 2-Hexanone-d5	8.627	63	81093	47.173	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.340%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	49059	5.172	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.400%	
66) 1,2-Dichlorobenzene-d4	12.187	152	54948	5.453	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	109.000%	
Target Compounds						
10) 1,1,2-Trichloro-1,2,2-...	2.576	101	4119	0.434	ug/L	92
12) 1,1-Dichloroethene	2.576	96	8252	0.951	ug/L	# 1
22) cis-1,2-Dichloroethene	4.666	96	2898	0.291	ug/L	88
34) Trichloroethene	6.541	95	8576	0.815	ug/L	93
47) Tetrachloroethene	8.550	164	6757	0.808	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091624\
Data File : VU060797.D
Acq On : 16 Sep 2024 22:15
Operator : MD/SY
Sample : P3967-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AN4

Quant Time: Sep 17 01:58:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Sep 17 01:47:23 2024
Response via : Initial Calibration

