

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092022\
 Data File : VU050858.D
 Acq On : 21 Sep 2022 02:15
 Operator : JC/MD
 Sample : N4650-18
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXZ13

Quant Time: Sep 21 04:42:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 21 04:30:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	367548	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	367220	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	138538	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	170138	44.679	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	89.360%	
7) Chloroethane-d5	1.906	69	136665	50.392	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	100.780%	
11) 1,1-Dichloroethene-d2	2.565	63	207907	35.088	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	70.180%	
21) 2-Butanone-d5	4.620	46	296024	120.410	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	120.410%	
24) Chloroform-d	5.064	84	292250	48.972	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.940%	
26) 1,2-Dichloroethane-d4	5.700	65	189924	51.641	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	103.280%	
32) Benzene-d6	5.726	84	582035	48.120	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.240%	
36) 1,2-Dichloropropane-d6	6.690	67	203542	49.854	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	99.700%	
41) Toluene-d8	7.896	98	464476	43.931	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.860%	
43) trans-1,3-Dichloroprop...	8.179	79	76362	42.994	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	85.980%	
47) 2-Hexanone-d5	8.633	63	174429	103.672	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	103.670%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	302304	48.566	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	97.140%	
66) 1,2-Dichlorobenzene-d4	12.192	152	158091	54.502	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	109.000%	
Target Compounds						
					Qvalue	
13) Acetone	2.630	43	5729	2.752	ug/L	97
19) 1,1-Dichloroethane	3.874	63	4815	0.823	ug/L	96
20) cis-1,2-Dichloroethene	4.668	96	5317	1.656	ug/L	98
30) 1,1,1-Trichloroethane	5.318	97	4180	0.856	ug/L #	35

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092022\
Data File : VU050858.D
Acq On : 21 Sep 2022 02:15
Operator : JC/MD
Sample : N4650-18
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXZ13

Quant Time: Sep 21 04:42:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 21 04:30:12 2022
Response via : Initial Calibration

