

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092922\
 Data File : VU051002.D
 Acq On : 30 Sep 2022 00:45
 Operator : JC/MD
 Sample : VU0929WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK191

Quant Time: Sep 30 02:30:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 30 02:29:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	324299	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	323354	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	112360	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	155049	47.176	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	94.360%	
7) Chloroethane-d5	1.909	69	191846	62.839	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	125.680%	
11) 1,1-Dichloroethene-d2	2.565	63	177461	36.553	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	73.100%	
21) 2-Butanone-d5	4.626	46	208250	97.695	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	97.690%	
24) Chloroform-d	5.060	84	256271	48.398	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.800%	
26) 1,2-Dichloroethane-d4	5.700	65	165766	50.577	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.160%	
32) Benzene-d6	5.726	84	510533	53.916	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	107.840%	
36) 1,2-Dichloropropane-d6	6.687	67	169162	55.055	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	110.100%	
41) Toluene-d8	7.896	98	423099	48.362	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	96.720%	
43) trans-1,3-Dichloroprop...	8.179	79	66648	45.477	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.960%	
47) 2-Hexanone-d5	8.636	63	111638	89.982	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	89.980%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	244792	45.585	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	91.180%	
66) 1,2-Dichlorobenzene-d4	12.192	152	132860	55.351	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	110.700%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092922\
Data File : VU051002.D
Acq On : 30 Sep 2022 00:45
Operator : JC/MD
Sample : VU0929WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK191

Quant Time: Sep 30 02:30:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 30 02:29:24 2022
Response via : Initial Calibration

