

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092922\
 Data File : VU051013.D
 Acq On : 30 Sep 2022 05:03
 Operator : JC/MD
 Sample : N4880-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGSJ1

Quant Time: Sep 30 05:54:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 30 02:29:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	296068	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	311803	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	130365	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	127341	42.440	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	84.880%		
7) Chloroethane-d5	1.906	69	116605	41.835	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	83.680%		
11) 1,1-Dichloroethene-d2	2.565	63	151817	34.253	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	68.500%		
21) 2-Butanone-d5	4.620	46	183666	94.377	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	94.380%		
24) Chloroform-d	5.060	84	226381	46.829	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	93.660%		
26) 1,2-Dichloroethane-d4	5.700	65	145911	48.764	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.520%		
32) Benzene-d6	5.726	84	449291	49.206	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.420%		
36) 1,2-Dichloropropane-d6	6.687	67	151084	50.993	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	101.980%		
41) Toluene-d8	7.896	98	361725	42.878	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	85.760%		
43) trans-1,3-Dichloroprop...	8.179	79	57111	40.413	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	80.820%		
47) 2-Hexanone-d5	8.632	63	96721	80.847	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	80.850%		
56) 1,1,2,2-Tetrachloroeth...	10.754	84	284467	54.936	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	109.880%		
66) 1,2-Dichlorobenzene-d4	12.192	152	146810	52.715	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	105.440%		
Target Compounds						Qvalue
13) Acetone	2.629	43	4224	2.718	ug/L	91
14) Carbon disulfide	2.793	76	5651	0.905	ug/L	100
46) Tetrachloroethene	8.552	164	7038	3.772	ug/L	95
64) 1,3-Dichlorobenzene	11.748	146	3598	0.824	ug/L	90
65) 1,4-Dichlorobenzene	11.832	146	4790	1.098	ug/L	90
67) 1,2-Dichlorobenzene	12.211	146	13931	3.106	ug/L	96
70) 1,2,4-trichlorobenzene	13.841	180	4059	1.676	ug/L #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092922\
Data File : VU051013.D
Acq On : 30 Sep 2022 05:03
Operator : JC/MD
Sample : N4880-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGSJ1

Quant Time: Sep 30 05:54:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 30 02:29:24 2022
Response via : Initial Calibration

