

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100422\
 Data File : VU051111.D
 Acq On : 04 Oct 2022 17:08
 Operator : JC/MD
 Sample : N4816-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 38025-100220

Quant Time: Oct 05 02:49:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U100422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Oct 05 02:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.115	96	47870	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.632	95	13490	0.858	ug/l	0.00
Spiked Amount 1.000			Recovery	=	86.000%	
68) 1,2-Dichlorobenzene-d4	12.192	152	14609	0.835	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.000%	
Target Compounds						
27) Chloroform	5.089	83	789904	28.621	ug/l	98
44) Bromodichloromethane	7.105	83	759125	38.337	ug/l	98
55) Dibromochloromethane	8.809	129	331087	24.692	ug/l	99
67) Bromoform	10.288	173	387259	50.037	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100422\
Data File : VU051111.D
Acq On : 04 Oct 2022 17:08
Operator : JC/MD
Sample : N4816-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
38025-100220

Quant Time: Oct 05 02:49:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U100422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Oct 05 02:45:21 2022
Response via : Initial Calibration

