

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100522\
 Data File : VU051121.D
 Acq On : 05 Oct 2022 12:52
 Operator : JC/MD
 Sample : N4853-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW826

Quant Time: Oct 06 03:03:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 06 02:55:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.246	114	350015	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	368951	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	119028	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	156225	44.042	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	88.080%	
7) Chloroethane-d5	1.896	69	139988	42.484	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	84.960%	
11) 1,1-Dichloroethene-d2	2.562	63	178862	34.135	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	68.280%	
21) 2-Butanone-d5	4.620	46	225918	98.196	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	98.200%	
24) Chloroform-d	5.060	84	288547	50.489	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.980%	
26) 1,2-Dichloroethane-d4	5.700	65	187324	52.955	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	105.920%	
32) Benzene-d6	5.726	84	560205	51.850	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	103.700%	
36) 1,2-Dichloropropane-d6	6.687	67	196639	56.088	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	112.180%	
41) Toluene-d8	7.896	98	434472	43.524	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.040%	
43) trans-1,3-Dichloroprop...	8.176	79	63300	37.854	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	75.700%	
47) 2-Hexanone-d5	8.632	63	107531	75.960	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	75.960%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	300840	49.099	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	98.200%	
66) 1,2-Dichlorobenzene-d4	12.192	152	140281	55.169	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	110.340%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100522\
Data File : VU051121.D
Acq On : 05 Oct 2022 12:52
Operator : JC/MD
Sample : N4853-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW826

Quant Time: Oct 06 03:03:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 06 02:55:50 2022
Response via : Initial Calibration

