




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100522\
 Data File : VU051129.D
 Acq On : 05 Oct 2022 16:23
 Operator : JC/MD
 Sample : N4858-01ME
 Misc : 5.47g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGSE8ME

Manual IntegrationsAPPROVED

Quant Time: Oct 06 03:04:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 06 02:55:50 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 10/06/2022
 Supervised By :Mahesh Dadoda 10/08/2022



TIC: VU051129.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

2.610min (+ 0.042) 0.01 ug/L

response 44

Ion	Exp%	Act%
63.00	100.00	100.00
98.00	75.40	0.00#
65.00	24.10	0.00#
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100522\
 Data File : VU051129.D
 Acq On : 05 Oct 2022 16:23
 Operator : JC/MD
 Sample : N4858-01ME
 Misc : 5.47g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGSE8ME

Manual IntegrationsAPPROVED

Quant Time: Oct 06 03:04:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 06 02:55:50 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 10/06/2022
 Supervised By :Mahesh Dadoda 10/08/2022



TIC: VU051129.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

2.523min (-0.045) 29.60 ug/L m

response 190021

Ion	Exp%	Act%
63.00	100.00	100.00
98.00	75.40	0.00#
65.00	24.10	0.00#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100522\
 Data File : VU051129.D
 Acq On : 05 Oct 2022 16:23
 Operator : JC/MD
 Sample : N4858-01ME
 Misc : 5.47g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGSE8ME

Manual IntegrationsAPPROVED

Quant Time: Oct 06 03:04:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 06 02:55:50 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 10/06/2022
 Supervised By :Mahesh Dadoda 10/08/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	428779	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	437312	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	179565	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	129298	29.755	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	59.500%#	
7) Chloroethane-d5	1.867	69	62711	15.536	ug/L	-0.04
Spiked Amount 50.000	Range 70	- 130	Recovery	=	31.080%#	
11) 1,1-Dichloroethene-d2	2.523	63	190021m	29.603	ug/L	-0.04
Spiked Amount 50.000	Range 60	- 125	Recovery	=	59.200%#	
21) 2-Butanone-d5	4.636	46	231503	82.140	ug/L	0.01
Spiked Amount 100.000	Range 40	- 130	Recovery	=	82.140%	
24) Chloroform-d	5.057	84	293487	41.920	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	83.840%	
26) 1,2-Dichloroethane-d4	5.697	65	191758	44.251	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	88.500%	
32) Benzene-d6	5.719	84	587388	45.868	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	91.740%	
36) 1,2-Dichloropropane-d6	6.687	67	203664	49.011	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	98.020%	
41) Toluene-d8	7.896	98	496053	41.925	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	83.860%	
43) trans-1,3-Dichloroprop...	8.182	79	78371	39.541	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	79.080%	
47) 2-Hexanone-d5	8.649	63	147864	88.124	ug/L	0.02
Spiked Amount 100.000	Range 45	- 130	Recovery	=	88.120%	
56) 1,1,2,2-Tetrachloroeth...	10.761	84	326080	44.899	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	89.800%	
66) 1,2-Dichlorobenzene-d4	12.195	152	182142	47.482	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	94.960%	
Target Compounds						
16) Methylene chloride	3.028	84	5193	1.183	ug/L	96
46) Tetrachloroethene	8.549	164	140355	53.633	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100522\
Data File : VU051129.D
Acq On : 05 Oct 2022 16:23
Operator : JC/MD
Sample : N4858-01ME
Misc : 5.47g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGSE8ME

Manual IntegrationsAPPROVED

Quant Time: Oct 06 03:04:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 06 02:55:50 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 10/06/2022
Supervised By :Mahesh Dadoda 10/08/2022

