

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101124\
 Data File : VU061116.D
 Acq On : 11 Oct 2024 14:13
 Operator : MD/SY
 Sample : VIBLK148
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK148

Quant Time: Oct 14 04:19:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 11 06:46:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	168522	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	155991	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	75411	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	40554	3.366	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.400%
7) Chloroethane-d5	1.911	69	36275	4.022	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.400%
11) 1,1-Dichloroethene-d2	2.560	65	19332	3.637	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.800%
20) 2-Butanone-d5	4.624	46	135378	46.908	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.820%
24) Chloroform-d	5.052	84	88347	4.161	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.200%
26) 1,2-Dichloroethane-d4	5.692	65	48648	4.322	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.400%
32) Benzene-d6	5.718	84	174135	4.234	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.600%
36) 1,2-Dichloropropane-d6	6.682	67	59418	4.399	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.000%
41) Toluene-d8	7.891	98	159084	4.138	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	8.174	79	22788	3.946	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.000%
46) 2-Hexanone-d5	8.627	63	114428	47.867	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.740%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	43962	4.438	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	54987	4.560	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101124\
Data File : VU061116.D
Acq On : 11 Oct 2024 14:13
Operator : MD/SY
Sample : VIBLK148
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK148

Quant Time: Oct 14 04:19:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 11 06:46:55 2024
Response via : Initial Calibration

