

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101124\
 Data File : VU061121.D
 Acq On : 11 Oct 2024 16:16
 Operator : MD/SY
 Sample : P4389-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E27H0

Quant Time: Oct 14 04:19:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 11 06:46:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	159168	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	143415	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	68376	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.586	65	37601	3.304	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.000%
7) Chloroethane-d5	1.901	69	32724	3.842	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.800%
11) 1,1-Dichloroethene-d2	2.550	65	17901	3.566	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.400%
20) 2-Butanone-d5	4.701	46	117914	43.258	ug/L	0.08
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.520%
24) Chloroform-d	5.049	84	85952	4.286	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.800%
26) 1,2-Dichloroethane-d4	5.695	65	45569	4.287	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.800%
32) Benzene-d6	5.711	84	163626	4.327	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
36) 1,2-Dichloropropane-d6	6.682	67	57746	4.650	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.888	98	149467	4.229	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	8.174	79	21441	4.038	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.800%
46) 2-Hexanone-d5	8.634	63	106624	48.514	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.020%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	42421	4.658	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.200%
66) 1,2-Dichlorobenzene-d4	12.184	152	52229	4.777	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.600%
Target Compounds						
					Qvalue	
13) Acetone	2.743	43	22510	12.840	ug/L	72


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101124\
Data File : VU061121.D
Acq On : 11 Oct 2024 16:16
Operator : MD/SY
Sample : P4389-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E27H0

Quant Time: Oct 14 04:19:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 11 06:46:55 2024
Response via : Initial Calibration





#13
 Acetone
 Concen: 12.840 ug/L
 RT: 2.743 min Scan# 41
 Delta R.T. 0.109 min
 Lab File: VU061121.D
 Acq: 11 Oct 2024 16:16

Instrument :
 MSVOA_U
 ClientSampleId :
 E27H0

Tgt Ion: 43 Resp: 22510
 Ion Ratio Lower Upper
 43 100
 58 18.7 0.0 69.6

