

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101824\
 Data File : VU061156.D
 Acq On : 18 Oct 2024 10:26
 Operator : MD/SY
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 18 23:45:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U101724DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Oct 18 02:52:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.551	0.608	-10.3	100	0.00
3 t	Chloromethane	0.524	0.574	-9.5	103	0.00
4 Rt	Vinyl Chloride	0.538	0.603	-12.1	102	0.00
5 T	Bromomethane	0.338	0.369	-9.2	98	0.00
6 T	Chloroethane	0.361	0.419	-16.1	108	0.00
7 T	Trichlorofluoromethane	0.766	0.976	-27.4	118	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.381	0.422	-10.8	99	0.00
9 Rt	1,1-Dichloroethene	0.397	0.451	-13.6	102	0.00
10 t	Iodomethane	0.472	0.620	-31.4#	116	0.00
11 t	Allyl Chloride	0.511	0.562	-10.0	98	0.00
12 t	Acrylonitrile	0.080	0.090	-12.5	104	0.00
13 T	Acetone	0.104	0.114	-9.6	93	0.00
14 T	Carbon Disulfide	1.367	1.497	-9.5	100	0.00
15 RT	Methylene Chloride	0.513	0.509	0.8	98	0.00
16 RT	trans-1,2-Dichloroethene	0.465	0.511	-9.9	99	0.00
17 t	1,1-Dichloroethane	0.825	0.921	-11.6	101	0.00
18 T	2-Butanone	0.127	0.139	-9.4	95	0.00
19	Cyclohexane	0.739	0.804	-8.8	98	0.00
20	Methylcyclohexane	0.824	0.915	-11.0	97	0.00
21 T	2,2-Dichloropropane	0.761	0.819	-7.6	98	0.00
22 RT	cis-1,2-Dichloroethene	0.507	0.564	-11.2	100	0.00
23 t	Diethyl Ether	0.308	0.338	-9.7	104	0.00
24 t	tert-Butyl Alcohol	0.032	0.041	-28.1	132	0.00
25 t	Methyl tert-Butyl Ether	1.063	1.185	-11.5	102	0.00
26 t	Bromochloromethane	0.212	0.239	-12.7	104	0.00
27 t	Chloroform	0.865	0.970	-12.1	102	0.00
28 RT	1,1,1-Trichloroethane	0.763	0.847	-11.0	99	0.00
29 T	1,1-Dichloropropene	0.674	0.736	-9.2	97	0.00
30 RT	Carbon Tetrachloride	0.662	0.743	-12.2	99	0.00
31 t	Isopropyl Ether	1.177	1.324	-12.5	101	0.00
32	Ethyl-t-butyl ether	1.210	1.368	-13.1	102	0.00
33	Tert-Amyl methyl ether	1.089	1.263	-16.0	104	0.00
34 t	Propionitrile	0.030	0.036	-20.0	101	0.00
35 RT	Benzene	1.898	2.140	-12.8	101	0.00
36 RT	1,2-Dichloroethane	0.536	0.596	-11.2	100	0.00
37 RT	Trichloroethene	0.507	0.559	-10.3	99	0.00
38 Rt	1,2-Dichloropropane	0.481	0.541	-12.5	100	0.00
39 t	Methacrylonitrile	0.109	0.130	-19.3	100	0.00
40 t	Methyl acrylate	0.222	0.246	-10.8	103	0.00
41 t	Tetrahydrofuran	0.066	0.069	-4.5	102	0.00
42 t	1-Chlorobutane	0.865	0.955	-10.4	97	0.00
43 T	Dibromomethane	0.249	0.275	-10.4	101	0.00
44 T	Bromodichloromethane	0.606	0.683	-12.7	101	0.00
45 T	4-Methyl-2-Pentanone	0.250	0.275	-10.0	99	0.00
46 t	t-1,4-Dichloro-2-butene	0.100	0.108	-8.0	90	0.00
47 t	Methyl methacrylate	0.217	0.250	-15.2	101	0.00
48 t	Ethyl methacrylate	0.462	0.528	-14.3	101	0.00
49 Rt	Toluene	1.199	1.354	-12.9	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101824\
 Data File : VU061156.D
 Acq On : 18 Oct 2024 10:26
 Operator : MD/SY
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 18 23:45:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U101724DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Oct 18 02:52:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.584	0.671	-14.9	100	0.00
51 T	cis-1,3-Dichloropropene	0.717	0.821	-14.5	100	0.00
52 RT	1,1,2-Trichloroethane	0.336	0.376	-11.9	102	0.00
53 t	1,3-Dichloropropane	0.591	0.657	-11.2	100	0.00
54 t	2-Hexanone	0.206	0.220	-6.8	92	0.00
55 t	Dibromochloromethane	0.397	0.458	-15.4	101	0.00
56 T	1,2-Dibromoethane	0.323	0.360	-11.5	100	0.00
57 S	4-Bromofluorobenzene	0.531	0.544	-2.4	95	0.00
58 RT	Tetrachloroethene	0.430	0.472	-9.8	99	0.00
59 Rt	Chlorobenzene	1.292	1.439	-11.4	99	0.00
60 T	1,1,1,2-Tetrachloroethane	0.436	0.501	-14.9	102	0.00
61 t	Pentachloroethane	0.333	0.395	-18.6	101	0.00
62 t	Hexachloroethane	0.344	0.410	-19.2	101	0.00
63 Rt	Ethyl Benzene	2.314	2.593	-12.1	99	0.00
64 RT	m/p-Xylenes	0.881	0.990	-12.4	98	0.00
65 RT	o-Xylene	0.867	0.978	-12.8	99	0.00
66 RT	Styrene	1.349	1.564	-15.9	100	0.00
67 t	Bromoform	0.208	0.246	-18.3	101	0.00
68 S	1,2-Dichlorobenzene-d4	0.524	0.556	-6.1	98	0.00
69 T	Isopropylbenzene	2.064	2.321	-12.5	98	0.00
70 T	1,1,2,2-Tetrachloroethane	0.408	0.467	-14.5	103	0.00
71 T	1,2,3-Trichloropropane	0.333	0.358	-7.5	99	0.00
72 t	Bromobenzene	0.496	0.549	-10.7	98	0.00
73 t	n-propylbenzene	0.602	0.694	-15.3	99	0.00
74 t	2-Chlorotoluene	0.521	0.588	-12.9	99	0.00
75 t	1,3,5-Trimethylbenzene	1.861	2.128	-14.3	99	0.00
76 t	4-Chlorotoluene	0.529	0.600	-13.4	98	0.00
77 t	tert-Butylbenzene	1.839	2.103	-14.4	99	0.00
78 t	1,2,4-Trimethylbenzene	1.878	2.151	-14.5	99	0.00
79 t	sec-Butylbenzene	2.395	2.746	-14.7	98	0.00
80	Nitrobenzene	0.009	0.011	-22.2	95	0.00
81 t	p-Isopropyltoluene	1.983	2.284	-15.2	99	0.00
82 t	1,3-Dichlorobenzene	0.952	1.084	-13.9	99	0.00
83 Rt	1,4-Dichlorobenzene	0.949	1.088	-14.6	99	0.00
84 t	n-Butylbenzene	1.840	2.162	-17.5	98	0.00
85 Rt	1,2-Dichlorobenzene	0.890	1.014	-13.9	100	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.061	0.073	-19.7	98	0.00
87 Rt	1,2,4-Trichlorobenzene	0.509	0.620	-21.8	100	0.00
88 t	Hexachlorobutadiene	0.288	0.331	-14.9	99	0.00
89 t	Naphthalene	0.898	1.099	-22.4	103	0.00
90 t	1,2,3-Trichlorobenzene	0.457	0.544	-19.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0