

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102224\
 Data File : VU061178.D
 Acq On : 22 Oct 2024 10:05
 Operator : MD/SY
 Sample : VSTD0.0531
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.05031

Quant Time: Oct 23 02:22:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM102224.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 23 02:22:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.257	114	27950	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	35713	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	11505	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	1300	0.031	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	982	0.042	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.687	67	1090	0.049	ug/L	0.00
8) Toluene-d8	7.903	98	2963	0.059	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.745	84	819	0.043	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	1385	0.031	ug/L	83
6) Trichloroethene	6.544	95	1005	0.044	ug/L	94
9) 1,2-Dibromoethane	8.923	107	591	0.042	ug/L #	96
12) 1,2,3-Trichloropropane	10.814	75	617	0.046	ug/L #	78
13) 1,2-Dibromo-3-chloropr...	12.990	75	115	0.045	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102224\
Data File : VU061178.D
Acq On : 22 Oct 2024 10:05
Operator : MD/SY
Sample : VSTD0.0531
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.05031

Quant Time: Oct 23 02:22:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM102224.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 23 02:22:24 2024
Response via : Initial Calibration

