

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102224\
 Data File : VU061185.D
 Acq On : 22 Oct 2024 15:55
 Operator : MD/SY
 Sample : P4476-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 XA110

Quant Time: Oct 23 02:34:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM102224.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 23 02:31:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	26880	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	35912	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	11310	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	11795	0.478	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	96.000%		
4) 1,2-Dichloroethane-d4	5.685	65	8901	0.479	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	96.000%		
7) 1,2-Dichloropropane-d6	6.687	67	10408	0.477	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	96.000%		
8) Toluene-d8	7.887	98	29120	0.474	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	94.000%		
10) 1,1,2,2-Tetrachloroeth...	10.745	84	7715	0.472	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	94.000%		
Target Compounds						Qvalue
9) 1,2-Dibromoethane	8.924	107	8514	0.641	ug/L	99
12) 1,2,3-Trichloropropane	10.814	75	18275	1.506	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.990	75	3711	1.560	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102224\
Data File : VU061185.D
Acq On : 22 Oct 2024 15:55
Operator : MD/SY
Sample : P4476-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
XA110

Quant Time: Oct 23 02:34:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM102224.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 23 02:31:01 2024
Response via : Initial Calibration

