

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102224\
 Data File : VU061183.D
 Acq On : 22 Oct 2024 12:45
 Operator : MD/SY
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VICV036

Quant Time: Oct 23 02:34:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM102224.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 23 02:31:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.257	114	27089	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	35213	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	10828	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	12053	0.485	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	98.000%
4) 1,2-Dichloroethane-d4	5.685	65	9329	0.498	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
7) 1,2-Dichloropropane-d6	6.687	67	10648	0.498	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	100.000%
8) Toluene-d8	7.887	98	30208	0.502	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	8178	0.511	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.597	62	13832	0.508	ug/L	100
6) Trichloroethene	6.544	95	10767	0.522	ug/L	99
9) 1,2-Dibromoethane	8.923	107	6908	0.531	ug/L	99
12) 1,2,3-Trichloropropane	10.814	75	6343	0.546	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.990	75	1295	0.569	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102224\
Data File : VU061183.D
Acq On : 22 Oct 2024 12:45
Operator : MD/SY
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VICV036

Quant Time: Oct 23 02:34:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM102224.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 23 02:31:01 2024
Response via : Initial Calibration

