

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102424\
 Data File : VU061204.D
 Acq On : 24 Oct 2024 17:23
 Operator : MD/SY
 Sample : P4533-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCNX0

Quant Time: Oct 25 02:38:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 25 02:35:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	190197	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	173436	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	77289	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	55737	4.314	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.200%
7) Chloroethane-d5	1.907	69	58600	4.947	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.000%
11) 1,1-Dichloroethene-d2	2.560	65	25257	4.569	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.400%
20) 2-Butanone-d5	4.621	46	123535	47.094	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.180%
24) Chloroform-d	5.052	84	113283	4.619	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
26) 1,2-Dichloroethane-d4	5.695	65	59620	4.877	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.600%
32) Benzene-d6	5.718	84	238462	4.880	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
36) 1,2-Dichloropropane-d6	6.682	67	71792	4.886	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.800%
41) Toluene-d8	7.891	98	220329	4.822	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
43) trans-1,3-Dichloroprop...	8.174	79	28471	4.747	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.000%
46) 2-Hexanone-d5	8.627	63	119186	48.521	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.040%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	53685	4.832	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	69770	5.098	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102424\
Data File : VU061204.D
Acq On : 24 Oct 2024 17:23
Operator : MD/SY
Sample : P4533-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCNX0

Quant Time: Oct 25 02:38:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 25 02:35:03 2024
Response via : Initial Calibration

