

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102424\
 Data File : VU061196.D
 Acq On : 24 Oct 2024 10:50
 Operator : MD/SY
 Sample : VU1024WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK064

Quant Time: Oct 25 02:36:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 25 02:35:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	177801	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	158433	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	69548	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	59350	4.914	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.200%
7) Chloroethane-d5	1.907	69	56003	5.058	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.200%
11) 1,1-Dichloroethene-d2	2.560	65	26119	5.054	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	101.000%
20) 2-Butanone-d5	4.624	46	130643	53.276	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.560%
24) Chloroform-d	5.055	84	115475	5.037	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.800%
26) 1,2-Dichloroethane-d4	5.695	65	61131	5.350	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.000%
32) Benzene-d6	5.718	84	240786	5.394	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.800%
36) 1,2-Dichloropropane-d6	6.682	67	71799	5.350	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.000%
41) Toluene-d8	7.891	98	220969	5.294	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.800%
43) trans-1,3-Dichloroprop...	8.174	79	28904	5.276	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.600%
46) 2-Hexanone-d5	8.627	63	122909	54.775	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.560%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	54741	5.394	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	66313	5.384	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102424\
Data File : VU061196.D
Acq On : 24 Oct 2024 10:50
Operator : MD/SY
Sample : VU1024WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK064

Quant Time: Oct 25 02:36:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 25 02:35:03 2024
Response via : Initial Calibration

