

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102524\
 Data File : VU061234.D
 Acq On : 25 Oct 2024 18:01
 Operator : MD/SY
 Sample : P4536-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B93

Quant Time: Oct 26 00:35:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 26 00:29:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	186025	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	165616	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	73087	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	46491	3.679	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.600%
7) Chloroethane-d5	1.911	69	43842	3.784	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.600%
11) 1,1-Dichloroethene-d2	2.560	65	22057	4.080	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.600%
20) 2-Butanone-d5	4.628	46	114446	44.607	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.220%
24) Chloroform-d	5.055	84	103895	4.331	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
26) 1,2-Dichloroethane-d4	5.695	65	53432	4.469	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
32) Benzene-d6	5.718	84	214760	4.602	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
36) 1,2-Dichloropropane-d6	6.682	67	64829	4.621	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.891	98	195352	4.477	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
43) trans-1,3-Dichloroprop...	8.174	79	23870	4.168	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.400%
46) 2-Hexanone-d5	8.627	63	101072	43.090	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.180%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	47131	4.442	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	62785	4.851	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102524\
Data File : VU061234.D
Acq On : 25 Oct 2024 18:01
Operator : MD/SY
Sample : P4536-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B93

Quant Time: Oct 26 00:35:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 26 00:29:08 2024
Response via : Initial Calibration

