

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102524\
 Data File : VU061231.D
 Acq On : 25 Oct 2024 16:47
 Operator : MD/SY
 Sample : P4533-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCNW3

Quant Time: Oct 26 00:33:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 26 00:29:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	195502	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	171006	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	77285	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	50245	3.783	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.600%
7) Chloroethane-d5	1.907	69	50071	4.113	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.200%
11) 1,1-Dichloroethene-d2	2.560	65	22942	4.038	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.800%
20) 2-Butanone-d5	4.624	46	126075	46.758	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.520%
24) Chloroform-d	5.052	84	108715	4.313	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
26) 1,2-Dichloroethane-d4	5.692	65	58122	4.626	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	5.718	84	230359	4.781	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.600%
36) 1,2-Dichloropropane-d6	6.6					

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102524\
Data File : VU061231.D
Acq On : 25 Oct 2024 16:47
Operator : MD/SY
Sample : P4533-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCNW3

Quant Time: Oct 26 00:33:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 26 00:29:08 2024
Response via : Initial Calibration

