

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102722\
 Data File : VU051539.D
 Acq On : 27 Oct 2022 14:57
 Operator : JC/MD
 Sample : VU1027WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK120

Quant Time: Oct 28 00:09:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 28 00:08:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	288124	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	282788	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	118391	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	79557	3.423	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.400%
7) Chloroethane-d5	1.909	69	78979	3.955	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.000%
11) 1,1-Dichloroethene-d2	2.565	65	31173	3.393	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.800%
20) 2-Butanone-d5	4.623	46	238892	46.686	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.380%
24) Chloroform-d	5.063	84	170141	3.927	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.600%
26) 1,2-Dichloroethane-d4	5.700	65	89860	4.021	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
32) Benzene-d6	5.726	84	340389	4.376	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
36) 1,2-Dichloropropane-d6	6.690	67	114681	4.341	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.800%
41) Toluene-d8	7.896	98	254457	3.656	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.200%
43) trans-1,3-Dichloroprop...	8.179	79	32529	3.748	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.000%
46) 2-Hexanone-d5	8.632	63	180936	61.849	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.700%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	92077	4.363	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.200%
66) 1,2-Dichlorobenzene-d4	12.192	152	91481	4.446	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102722\
Data File : VU051539.D
Acq On : 27 Oct 2022 14:57
Operator : JC/MD
Sample : VU1027WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK120

Quant Time: Oct 28 00:09:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 28 00:08:00 2022
Response via : Initial Calibration

