

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102824\
 Data File : VU061261.D
 Acq On : 28 Oct 2024 10:19
 Operator : MD/SY
 Sample : VU1028WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK068

Quant Time: Oct 29 04:46:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 29 04:45:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	192443	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	175634	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	78688	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	45398	3.473	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.400%
7) Chloroethane-d5	1.907	69	40110	3.347	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	67.000%
11) 1,1-Dichloroethene-d2	2.557	65	19849	3.549	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.000%
20) 2-Butanone-d5	4.631	46	122736	46.243	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.480%
24) Chloroform-d	5.052	84	99749	4.020	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.400%
26) 1,2-Dichloroethane-d4	5.692	65	51627	4.174	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.400%
32) Benzene-d6	5.717	84	202238	4.087	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.800%
36) 1,2-Dichloropropane-d6	6.682	67	62469	4.199	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.000%
41) Toluene-d8	7.891	98	176146	3.807	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.200%
43) trans-1,3-Dichloroprop...	8.174	79	24018	3.955	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.000%
46) 2-Hexanone-d5	8.627	63	103356	41.550	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.100%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	46636	4.145	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	56973	4.089	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102824\
Data File : VU061261.D
Acq On : 28 Oct 2024 10:19
Operator : MD/SY
Sample : VU1028WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK068

Quant Time: Oct 29 04:46:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 29 04:45:10 2024
Response via : Initial Calibration

