

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102824\
 Data File : VU061267.D
 Acq On : 28 Oct 2024 15:16
 Operator : MD/SY
 Sample : P4533-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Oct 29 04:48:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 29 04:45:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	177576	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	159509	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	72798	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	45656	3.785	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.600%
7) Chloroethane-d5	1.911	69	43858	3.966	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.400%
11) 1,1-Dichloroethene-d2	2.560	65	20240	3.922	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.400%
20) 2-Butanone-d5	4.627	46	125061	51.064	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.120%
24) Chloroform-d	5.055	84	101791	4.446	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
26) 1,2-Dichloroethane-d4	5.692	65	53900	4.723	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
32) Benzene-d6	5.717	84	201032	4.473	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
36) 1,2-Dichloropropane-d6	6.682	67	63260	4.682	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.600%
41) Toluene-d8	7.891	98	174019	4.141	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	8.174	79	24114	4.372	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.400%
46) 2-Hexanone-d5	8.627	63	110490	48.908	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.820%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	50304	4.923	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	58745	4.557	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102824\
Data File : VU061267.D
Acq On : 28 Oct 2024 15:16
Operator : MD/SY
Sample : P4533-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Oct 29 04:48:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 29 04:45:10 2024
Response via : Initial Calibration

